Day Surgery 19/05/2026

Analysis	S/N 413 Left - 7.3 19/05/2026 11:30 SEQ Medical Pty	S/N 413 Right - 7.3 19/05/2026 11:30 SEQ Medical Pty
Total Viable Count (CFU/100mL)	99	99
Pseudomonas aeruginosa (CFU/100mL)	<1	<1
Microorganism ID	Acinetobacter pittii	Acinetobacter pittii

Table 7.3

Final Rinse Water - Washer-Disinfectors in Accordance with IOS 15883-4 for Thermolabile Endoscopes

Substance	Specification
Pseudomonas aeruginosa (CFU/100mL)	≤ 1
Total Viable Count (CFU/100mL)	≤ 10.1
Microorganism ID	—